

Modello Laser	E-400
Serial Number RF	4PA11370194
Serial Number Sorgente	ET53633
Sample Rate:	2000
Sample Period:	00.00.3
Total Duration:	0.11.23

Data	08/03/2013 14.21
Sonda	0164L12R
Frequenza	50 KHz
Duty	60 %
Valore Minimo	511 W
Valore Medio	521 W
Valore Massimo	539 W
Deviazione Standard	4.5 W
Stabilità	1.74% +/-

