

Modello Laser	G-150
Serial Number RF	-
Serial Number Sorgente	GB45560
Sample Rate:	2000
Sample Period:	00:00.3
Total Duration:	0:11:20

Data	05/02/2013 10:36
Sonda	0164L12R
Frequenza	50 KHz
Duty	60 %
Valore Minimo	181 W
Valore Medio	186 W
Valore Massimo	191 W
Deviazione Standard	2.8 W
Stabilità	3.01% +/-

