

Modello Laser	K-250
Serial Number RF	135MP1301003
Serial Number Sorgente	HB47303
Sample Rate:	2000
Sample Period:	00:00,3
Total Duration:	0:11:19

Data	19/11/2013 16:13
Sonda	0163L12R
Frequenza	1 KHz
Duty	60 %
Valore Minimo	331 W
Valore Medio	341 W
Valore Massimo	349 W
Deviazione Standard	4,6 W
Stabilità	2,69% +/-

