

Modello Laser	<b>K-250</b>
Serial Number RF	<b>135MQ1202008</b>
Serial Number Sorgente	<b>HB47154</b>
Sample Rate:	<b>2000</b>
Sample Period:	<b>00:00,3</b>
Total Duration:	<b>0:11:20</b>

Data	<b>18/07/2013 10:11</b>
Sonda	<b>0164L12R</b>
Frequenza	<b>50 KHz</b>
Duty	<b>60 %</b>
Valore Minimo	<b>305 W</b>
Valore Medio	<b>317 W</b>
Valore Massimo	<b>328 W</b>
Deviazione Standard	<b>6,0 W</b>
Stabilità	<b>3,76% +/-</b>

